


Linearity Test

	Weight Applied	As Found				As Left			
		Reading	Error	Allowable Error	✓	Reading	Error	Allowable Error	✓
Zero 1	0 lb	0 lb	0 lb	20 lb	✓	N/A	N/A	N/A	N/A
2	4,400 lb	4,400 lb	0 lb	20 lb	✓	N/A	N/A	N/A	N/A
3	8,820 lb	8,820 lb	0 lb	20 lb	✓	N/A	N/A	N/A	N/A
4	13,220 lb	13,220 lb	0 lb	40 lb	✓	N/A	N/A	N/A	N/A
5	17,640 lb	17,640 lb	0 lb	40 lb	✓	N/A	N/A	N/A	N/A
Max (x)	22,060 lb	22,060 lb	0 lb	40 lb	✓	N/A	N/A	N/A	N/A
5	17,640 lb	17,640 lb	0 lb	40 lb	✓	N/A	N/A	N/A	N/A
4	13,220 lb	13,220 lb	0 lb	40 lb	✓	N/A	N/A	N/A	N/A
3	8,820 lb	8,820 lb	0 lb	20 lb	✓	N/A	N/A	N/A	N/A
2	4,400 lb	4,400 lb	0 lb	20 lb	✓	N/A	N/A	N/A	N/A
Zero 1	0 lb	0 lb	0 lb	20 lb	✓	N/A	N/A	N/A	N/A

Strain Load Test

	Weight of Empty Truck	Amount of Test Weights	Indication of Truck and Weights	Error on Test Weights Only	Allowable Error	✓
1	29,780 lb	22,060 lb	51,840 lb	0 lb	40 lb	✓

Shift Test #1 (Single Platform Sections Only)



	Test Load: 51,840 lb						As Found						As Left					
	Far		Middle		Near		Far		Middle		Near		Far		Middle		Near	
	Reading	Error	Reading	Error	Reading	Error	Reading	Error	Reading	Error	Reading	Error	Reading	Error	Reading	Error	Reading	Error
1			51,840 lb	0 lb					N/A	N/A					N/A	N/A		
2			51,840 lb	0 lb					N/A	N/A					N/A	N/A		
3			51,840 lb	0 lb					N/A	N/A					N/A	N/A		
4			51,840 lb	0 lb					N/A	N/A					N/A	N/A		
5			51,840 lb	0 lb					N/A	N/A					N/A	N/A		
6			51,840 lb	0 lb					N/A	N/A					N/A	N/A		

	As Found	As Left
Range Of Results	0 lb	N/A
Allowable Error	80 lb	N/A
Within Tolerances	✓	N/A

Reference Weights

Weight Set	Traceability Number	Class ASTM/OIML	Calibration Date	Calibration Due Date
A1 - A20	1408974	ASTM 5	17-03-2021	17-03-2022

Contact Details

Customer Contact	Position	Phone	Email
Dan Theriault			

This document is issued to record completion of the work performed by METTLER TOLEDO on the subject device in accordance with agreed standards. It does not guarantee the continued performance of the subject device. Any measurements recorded are based on the subject device's performance at a given time as tested by METTLER TOLEDO and, except where explicitly stated otherwise, do not express an opinion as to the sufficiency of any customer designed procedures used to test the device. This document is not a warranty, either implied or express. METTLER TOLEDO expressly disclaims any liability arising from the use of the information in this document for any purpose other than as specified herein.