


Linearity Test

	Weight Applied	As Found				As Left			
		Reading	Error	Allowable Error	✓	Reading	Error	Allowable Error	✓
Zero 1	0 lb	0 lb	0 lb	20 lb	✓	0 lb	0 lb	20 lb	✓
2	4,400 lb	4,400 lb	0 lb	20 lb	✓	4,400 lb	0 lb	20 lb	✓
3	8,820 lb	8,820 lb	0 lb	20 lb	✓	8,820 lb	0 lb	20 lb	✓
4	13,220 lb	13,220 lb	0 lb	40 lb	✓	13,220 lb	0 lb	40 lb	✓
5	17,640 lb	17,640 lb	0 lb	40 lb	✓	17,640 lb	0 lb	40 lb	✓
Max (x)	22,060 lb	22,060 lb	0 lb	40 lb	✓	22,060 lb	0 lb	40 lb	✓
5	17,640 lb	17,640 lb	0 lb	40 lb	✓	17,640 lb	0 lb	40 lb	✓
4	13,220 lb	13,220 lb	0 lb	40 lb	✓	13,220 lb	0 lb	40 lb	✓
3	8,820 lb	8,820 lb	0 lb	20 lb	✓	8,820 lb	0 lb	20 lb	✓
2	4,400 lb	4,400 lb	0 lb	20 lb	✓	4,400 lb	0 lb	20 lb	✓
Zero 1	0 lb	0 lb	0 lb	20 lb	✓	0 lb	0 lb	20 lb	✓

Shift Test #1 (Single Platform Sections Only)



	Test Load: 51,680 lb											
	As Found						As Left					
	Far		Middle		Near		Far		Middle		Near	
	Reading	Error	Reading	Error	Reading	Error	Reading	Error	Reading	Error	Reading	Error
1			51,660 lb	-20 lb					51,680 lb	0 lb		
2			51,680 lb	0 lb					51,680 lb	0 lb		
3			51,680 lb	0 lb					51,680 lb	0 lb		
4			51,700 lb	20 lb					51,680 lb	0 lb		
5			51,740 lb	60 lb					51,700 lb	20 lb		
6			51,720 lb	40 lb					51,680 lb	0 lb		

	As Found	As Left
Range Of Results	80 lb	20 lb
Allowable Error	80 lb	80 lb
Within Tolerances	✓	✓

Reference Weights

Weight Set	Traceability Number	Class ASTM/OIML	Calibration Date	Calibration Due Date
A1_A20	1418564	ASTM 5	31-03-2022	31-03-2023

Contact Details

Customer Contact	Position	Phone	Email
Dan T			

This document is issued to record completion of the work performed by METTLER TOLEDO on the subject device in accordance with agreed standards. It does not guarantee the continued performance of the subject device. Any measurements recorded are based on the subject device's performance at a given time as tested by METTLER TOLEDO and, except where explicitly stated otherwise, do not express an opinion as to the sufficiency of any customer designed procedures used to test the device. This document is not a warranty, either implied or express. METTLER TOLEDO expressly disclaims any liability arising from the use of the information in this document for any purpose other than as specified herein.